

## ANNEXURE A

**SCHEDULE OF ACCREDITATION**

## TEMPERATURE METROLOGY

Accreditation Number: CAL 095-03-00

<b>Permanent Address of Laboratory:</b> United Scientific (Pty) Ltd Unit 2B 16 Wessel Geldenhuys Street Brackenfell Industrial Cape Town 7460		<b>Technical Signatory:</b> Mr GB Mamaila Mr A Barnes		
<b>Postal Address:</b> PO Box 505 Goodwood Cape Town 7459		<b>Nominated Representative:</b> Mrs B Nhlapo		
Tel: 021 592 5240 Cell: 082 686 1470 E-mail: <a href="mailto:sanas@united-scientific.co.za">sanas@united-scientific.co.za</a>		Issue No.: 04 Date of Issue: 26 March 2026 Expiry Date: 11 April 2027		
ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT	RANGE OF MEASURED QUANTITY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	METHOD/ PROCEDURE
1	<b>THERMOMETRY</b>			
1.3	<b>Thermocouples</b>			
1.3.1	Liquid-in glass Thermometers	- 20 °C to 100 °C	0,5 K	Calibration by comparison with a reference thermometer in a bath, drywell, or furnace.
1.3.2	Digital Thermometers	- 25 °C to 200 °C	0,5 K	
1.4	<b>Reference Temperature Sources</b>			
1.4.1	Ice Point reference	0,00 °C	0,3 K	Prepared in a thermally insulated flask using distilled water and ice.
1.5	<b>Temperature Measuring and Recording</b>			
1.5.2	Data Loggers (Internal sensor)	- 25°C to 125°C	0,5 K	Calibration by comparison with a reference thermometer in a chamber, bath drywell or furnace.
	Data Logger (External sensor)	-25°C to 200 °C	0,5 K	

Original Date of Accreditation: 18 August 2023

Page 1 of 2

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor  $k = 2$ , corresponding to a confidence level of approximately 95%



**Accreditation Manager**

# ANNEXURE A

Accreditation No.: CAL 095-03-00  
 Date of Issue: 26 March 2026  
 Expiry Date: 11 April 2027

ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT	RANGE OF MEASURED QUANTITY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	METHOD / PROCEDURE
<b>4</b>	<b>TEMPERATURE INSTALLATIONS AND DEVICES</b>			
<b>4.1</b>	<b>Iso-thermal Media evaluation (Multi location over time monitoring)</b>			
4.1.1	Steam Sterilizers (Temperature, Pressure, Time)	20 °C to 125 °C 0 to 400 kPa (abs) 0 to 120 minutes	2,0 K 30 kPa 10 s	Calibration by temperature mapping over time using reference thermometers and/or loggers including calibration of the timing and pressure indicating device.
4.1.2	Environmental Chambers	- 80 °C to 10 °C 10 °C to 60 °C 60 °C to 150 °C	2,0 K 2,0 K 4,4 K	Calibration by temperature mapping over time using reference thermometers and/or loggers.
4.1.3	Furnace, Ovens			
4.1.4	Fridges, Freezers			
4.1.5	Incubators			
4.1.6	Liquid baths			

Original Date of Accreditation: 18 August 2023

Page 2 of 2

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor  $k = 2$ , corresponding to a confidence level of approximately 95%

ISSUED BY THE SOUTH AFRICAN NATIONAL ACCREDITATION SYSTEM



**Accreditation Manager**